

18

CORRIGENDUM

This is with reference to Tender Notice published online on 22.06.2019. It is to be informed to all the vendors that there has been an amendment in the Specifications in the Tender No. **CBMR/ PUR/585/2019-20 due on 20.07.2019** for purchase of HPLC.

Hence the following amendments in the Specifications in the Tender No CBMR/ PUR/585/2019-20 due on 20.07.2019 for purchase of HPLC may please be noted by all the applicants.

Amendment for HPLC Specifications (Tender No. CBMR/ PUR/585/2019-20 due on 20.07.2019)

The followings are the amendment of the specifications...

Earlier specification	Changed specification
Photo-diode array detector <ul style="list-style-type: none">• Detection type: 1024 photo diodes or more	Photo-diode array detector <ul style="list-style-type: none">• Detection type: 512/1024 photo diodes

(Prof. Raja Roy)
DIRECTOR

(Prof. RAJA ROY)
Director
Centre of Biomedical Research
SGPGIMS Campus Raibareilly Road
Lucknow-226014 (INDIA)